

Calibration Certificate

Control Number *****-001

(1/3)
Rev 1.05

Customer Name: *****

Issue Date 1-Jul-2014

Model: MA4701A
 Description: Power Sensor
 Manufacture: Anritsu Company
 Serial Number: *****
 Calibration Date: 1-Jul-2014

ANRITSU CUSTOMER SUPPORT CO.,LTD.
 Service Center
 5-1-1 Onna, Atsugi-shi, Kanagawa, 243-0032 Japan
 Certified by:

 Manager

We hereby certifies that this instrument was calibrated to following instruments/standards and conforms with the traceability chart.

This certificate consists of the used instruments/standards list and the traceability chart to show traceability from the calibrated measuring instrument to the national standard.

Certificate (or control) number and Calibration due date conforms when calibrated to each following instruments/standards.

SAMPLE

Traceability	Reg No.	Model	Description	Serial No.	Cycle	Cal. due date	Certificate No.
Measuring Instrument		MA4701A	Power Sensor	*****			
Shop Instruments	EA4-3385	MA4701A	Power Sensor	M78685	6	23-Dec-2014	14062313
Interlaboratory Standards	EA4-7364	8481A	Power Sensor	MY4109667	12	13-Nov-2014	JQA 150-34115
Shop Instruments	EA6-3290	560-97NF50-1	SWR Auto Tester	95462	12	29-Jul-2014	1307293290
Working Standards	EP8-0257	18N50	Air Line	504059	24	5-Mar-2015	*01 US00070535
Shop Instruments	EA6-6233	29A50-20	Offset Termination	008207	12	28-Feb-2014	1302286233
Shop Instruments	EM4-3743	34401A	DMM	US36045560	6	23-Aug-2013	13022001
Working Standards	EY3-0045	5700A	Calibrator	5985306	12	14-Nov-2013	A-PCU06J1375

Traceability	Reg No.	Model	Description	Serial No.	Cycle	Cal. due date	Certificate No.
Interlaboratory Standards	EM4-7185	3458A	DMM	MY4504012	12	24-Feb-2013	*02 120649
Working Standards	EM4-9344	3458A	DMM	MY4504012	12	6-Jul-2013	A-PCU06J1319
Interlaboratory Standards	EY3-6516	5700A	Calibrator	8555605	12	1-Dec-2012	JEMIC 011-115885-100
Shop Instruments	ES3-6773	37397C	Network Analyzer	010810	6	30-May-2013	1211306773
Interlaboratory Standards	EA6-0041	3669B-1	Verification Kit	89004	12	7-Nov-2013	*01 US00066284
Working Standards	EA6-6261	3654B	Calibration Kit	305003	12	29-Oct-2013	*01 US00066001
Interlaboratory Standards	EA6-7259	T1968-2	Air Line	0605001	24	2-Aug-2014	*01 US00062842
Shop Instruments	EA4-4005	MA2442A	Power Sensor	971926	6	3-Dec-2013	1306034005
Interlaboratory Standards	EA4-5088	8478B	Power Sensor	3318A25140	24	28-Jun-2014	NIST 37779839

SAMPLE

National standard

NIST:National Institute of Standards
and Technology

Other calibration laboratory

*01:Anritsu Company

*02:Agilent Technologies

JEMIC:Japan Electric Meters Inspection
Corporation

JQA:Japan Quality Assurance
Organization

SAMPLE