

EU DECLARATION OF CONFORMITY

Manufacturer's Name: ANRITSU COMPANY

Manufacturer's Address: 490 Jarvis Drive
Morgan Hill, CA 95037-2809
USA

declares that the product specified below:

Product Name: Calibration Component

Model Number: OSLK50, OSLKF50, OSLN50A-18, OSLN50A-8, OSLNF50A-18, OSLNF50A-8, TOSLK50A-20, TOSLK50A-40, TOSLKF50A-20, TOSLKF50A-40, TOSLN50A-18, TOSLN50A-8, TOSLNF50A-18, TOSLNF50A-8, 1091-53-R, 1091-54-R, 1091-55-R, 1091-56-R, 22A50, 22KF50, 22N50, 22N75, 22NF50, 22NF75, 22V50, 22VF50, 24L50, 24LF50, 22K50, 23L50, 23LF50, 1091-172-R, 1091-81-R, 11N50B, 12N50-75B, 34NFNF50, 2000-1618-R, 2000-1619-R, V210, 2000-1724-R, 2000-1749-R

conforms to the requirement of:

EMC Directive: N/A
Low Voltage Directive: N/A
RoHS Directive: 2011/65/EU

Electromagnetic Compatibility: EN 61326-1:2013
Emissions: EN 55011:2009 +A1:2010 Group 1 Class A
Immunity: EN 61000-4-2:2009 4 kV CD, 8 kV AD
EN 61000-4-3:2006 +A2:2010 3 V/m
EN 61000-4-4:2012 0.5 kV S-L, 1 kV P-L
EN 61000-4-5:2014 0.5 kV L-L, 1 kV L-E
EN 61000-4-6:2014 3 V
EN 61000-4-8:2010 3 A/m
EN 61000-4-11: 2004 100% @ 20 ms
Electrical Safety: EN 61010-1:2010



Eric McLean, Corporate Quality Director

June 5, 2017

Date

European Contact: For Anritsu product CE information, contact Anritsu EMEA Limited, 200 Capability Green, Luton, Bedfordshire, LU1 3LU, England. (Telephone: +44 (0)1582 433200; Email: bert.francis@anritsu.com)