

# EU DECLARATION OF CONFORMITY

**Manufacturer's Name:** ANRITSU COMPANY

**Manufacturer's Address:** 490 Jarvis Drive  
Morgan Hill, CA 95037-2809  
USA

declares that the product specified below:

**Product Name:** Cable Assembly

**Model Number:** 15KKF50-1.5A, 15NNF50-1.5B, 16DD50-2.75-R, 16DD50-4.0-R, 2000-1371-R, 2000-1536-R, 2000-1537-R, 2000-1538-R, 2000-1539-R, 2000-1540-R, 2000-1541-R, 2000-1542-R, 2000-1543-R, 2000-1566-R, 2000-1593-R, 2000-1594-R, 2000-1605-R, 2000-1606-R, 2000-1626-R, 2000-1732-R, 2000-1783-R, 2000-1809-R, 2000-1816-R, 2000-1845-R, 2000-1846-R, 2000-1859-R, 2000-1888-R, 2100-28-R, 2100-29-R, 2100-30-R, 2100-31-R, 3-2000-1498, 3-806-116, 3-806-117, 3-806-152, 3-806-169, 806-16-R, 806-176-R, 806-177-R, 808-16-R, 808-17-R, 808-18-R, 808-19-R, 2100-1, 2100-2, 2100-4, 3-806-116, 3-806-117, K101F-R, SC7496, SC7543, SC8218, N120-6, NS120MF-6, 806-245-R, 800-441, 806-206-R, 806-209-R, 806-304-R, 806-396-R, 15NNF50-1.0B, 15NNF50-1.5C, 15NNF50-5.0C

conforms to the requirement of:

**EMC Directive:** 2014/30/EU

**Low Voltage Directive:** 2014/35/EU


**RoHS Directive:** 2011/65/EU

**Electromagnetic Compatibility:** EN 61326-1:2013

**Emissions:** EN 55011:2009 +A1:2010 Group 1 Class A

<b>Immunity:</b>	EN 61000-4-2:2009	4 kV CD, 8 kV AD
	EN 61000-4-3:2006 +A2:2010	3 V/m
	EN 61000-4-4:2012	0.5 kV S-L, 1 kV P-L
	EN 61000-4-5:2014	0.5 kV L-L, 1 kV L-E
	EN 61000-4-6:2014	3 V
	EN 61000-4-8:2010	3 A/m
	EN 61000-4-11:2004	100% @ 20 ms

**Electrical Safety:** EN 61010-1:2010

  
\_\_\_\_\_  
Mekonen Buzuayene, Test Engineering Manager

\_\_\_\_\_  
October 26, 2020  
Date

European Contact: For Anritsu product CE information, contact Anritsu EMEA Limited, 200 Capability Green,

Luton, Bedfordshire, LU1 3LU, England. (Telephone: +44 (0)1582 433200; Email: [bert.francis@anritsu.com](mailto:bert.francis@anritsu.com))