

EU DECLARATION OF CONFORMITY

Manufacturer's Name: ANRITSU COMPANY

Manufacturer's Address: 490 Jarvis Drive
Morgan Hill, CA 95037-2809
USA

declares that the product specified below:

Product Name: Field Master Pro

Model Number: MS2090A-0709, MS2090A-0714, MS2090A-0720, MS2090A-0726, MS2090A-0732
MS2090A-0744, MS2090A-0754, MS2090A-0031, MS2090A-0090, MS2090A-0103
MS2090A-0104, MS2090A-0105, MS2090A-0106, MS2090A-0124, MS2090A-0125
MS2090A-0126, MS2090A-0199, MS2090A-0331, MS2090A-0360, MS2090A-0883
MS2090A-0888, MS2090A-0889

conforms to the requirement of:

EMC Directive: 2014/30/EU + FCC ID: SQG-60SIPT
Low Voltage Directive: 2014/35/EU
RoHS Directive: 2011/65/EU + Amendment 2015/863


Electromagnetic Compatibility: EN 61326-1:2013

Emissions: EN 55011:2009 +A1:2010 Group 1 Class A

Immunity:

EN 61000-4-2:2009	4 kV CD, 8 kV AD
EN 61000-4-3:2006 +A2:2010	3 V/m
EN 61000-4-4:2012	0.5 kV S-L, 1 kV P-L
EN 61000-4-5:2014	0.5 kV L-L, 1 kV L-E
EN 61000-4-6:2014	3 V
EN 61000-4-8:2010	3 A/m
EN 61000-4-11:2004	100% @ 20 ms

Electrical Safety: EN 61010-1:2010


Mekonen Buzuayene, Test Engineering Manager

October 26, 2021
Date

European Contact: For Anritsu product CE information, contact Anritsu EMEA Limited, 200 Capability Green, Luton, Bedfordshire, LU1 3LU, England. (Telephone: +44 (0)1582 433200; Email: bert.francis@anritsu.com)